

Channel	Part No.	Volume Range (µl)	Volume Increments (µl)	Test Volume (µl)	Accuracy (±%)	Precision (≤%)	Tip Spacing (mm)
4	4743	10 - 300	0.5	30 150 300	4.00 2.00 1.60	1.20 0.60 0.35	9.0 - 33.0
	4744	50 - 1250	1	125 625 1250	6.00 2.40 1.20	1.10 0.50 0.30	9.0 - 33.0
6	4763	10 - 300	0.5	30 150 300	4.00 2.00 1.60	1.20 0.60 0.35	9.0 - 19.8
	4764	50 - 1250	1	125 625 1250	6.00 2.40 1.20	1.10 0.50 0.30	9.0 - 19.8
8	4721	0.5 - 12.5	0.01	1.25 6.25 12.5	10.00 4.00 2.00	6.00 1.60 0.80	4.5 - 14.0
	4726	2 - 50	0.05	5 25 50	5.00 2.50 1.50	2.50 0.60 0.40	4.5 - 14.0
	4722	5 - 125	0.1	12.5 62.5 125	3.75 2.50 1.60	1.50 0.70 0.35	4.5 - 14.0
	4723	10 - 300	0.5	30 150 300	4.00 2.00 1.60	1.20 0.60 0.35	9.0 - 14.0
	4724	50 - 1250	1	125 625 1250	6.00 2.40 1.20	1.10 0.50 0.30	9.0 - 14.0
12	4731	0.5 - 12.5	0.01	1.25 6.25 12.5	10.00 4.00 2.00	6.00 1.60 0.80	4.5 - 9.0
	4736	2 - 50	0.05	5 25 50	5.00 2.50 1.50	2.50 0.60 0.40	4.5 - 9.0
	4732	5 - 125	0.1	12.5 62.5 125	3.75 2.50 1.60	1.50 0.70 0.35	4.5 - 9.0

1. The multichannel specifications represent the performance of all channels of a pipette.
2. Precision = coefficient of variation